

## EN) DUROSIL L - DUROSIL S - PASTE HARDENER®

C Silikones  
Instructions For Use1. INTENDED USE  
C-silicone dental impression material

**2. PRODUCT DESCRIPTION**  
Durosil L: High viscosity condensation polysiloxane (base) recommended for the two-phase technique (double impression) in combination with Durosil S.  
Durosil S: Low-viscosity condensation polysiloxane (base) recommended for the two-phase technique (double step impression) in combination with Durosil L.

Paste Hardener: Catalyst for condensation polysiloxanes. To be used for the polymerisation of Durosil L and Durosil S bases only. The Paste Hardener catalyst can be supplied separately from the bases of the materials it is combined with for mixing.


**3. PACKAGING**  
• 900 ml can of Durosil L • 140 ml tube of Durosil S • 60 ml tube of Paste Hardener

**4. COMPOSITION**  
Durosil L: Polysiloxanes (hydroxy terminated), silica fillers, inorganic aluminum fillers, hydrocarbons, pigments, flavoring agents.  
Durosil S: Polysiloxanes (hydroxy terminated), silica fillers, surfactants, pigments, flavors (fruitmix).  
Paste Hardener: Organogel compound, alkoxysilanes, hydrocarbons, pigments, mint flavor.


**5. INSTRUCTIONS FOR USE**  
The devices are intended for use in the dental sector by trained and qualified professionals for the purpose of impression taking on dental patients from 3 years of age.

**Preparing the Devices:** step-by-step instructions


Durosil L & Paste Hardener:



Durosil S & Paste Hardener:



**Disinfecting and Cleaning the Device: General step-by-step Instructions:**



**6. CONTRAINDICATIONS**  
Do not use on patients with known hypersensitivity to any of the components. For more information on the allergens present, please contact President Dental GmbH.

**7. SIDE EFFECTS**  
Irritation, redness or signs of hypersensitivity may occur in case of allergy to any of the components.

**8. CLINICAL BENEFITS**

Impression material for negative reproduction of dental arches for use in diagnosis and treatment.

**9. STEP-BY-STEP INSTRUCTIONS**

**9.1 GENERAL PRECAUTIONS/WARNINGS:**

The product's instructions for use must be kept for the duration of its use. Use gloves, protective goggles, face mask, and suitable clothing. Use on previous occasions and in the event of re-use, avoid cross-contamination of the impression material with other unsterile/uncontaminated gloves when handling cartridges, tubes, tubes and accessories. Packs should be kept in environments that are free of microbial contamination. Store the product at temperatures between 5°C / 41°F and 27°C / 80°. and keep away from direct sunlight. Vapours with a characteristic alcohol odour may be released during mixing; do not inhale. Do not ingest. If ingested, seek immediate medical advice. Safety Data Sheet available subject to request. Dispose of the product responsibly. The product (material, polymerised material and impression) and packaging must be disposed of in compliance with applicable local regulations.

**9.2 GENERAL PRECAUTIONS/WARNINGS ON DEVICE PREPARATION:**

Do not overfill the impression tray, as this could ingestion and trigger the patient's emetic reflex. Dispense the strips of product trying to keep the diameter of the dispensed material the same as the outlet hole on the tube. Avoid mixing with the strip when dispensing. Do not use if the material has been previously and partially dispensed. Do not use after the expiry date indicated on the packaging. Do not use the product if the main batch details and expiry date are not indicated on the outer packaging. Do not use the product if the main packaging is damaged. The devices should be used within a professional dental environment (legally-certified public or private health facilities). Use the product at an ambient temperature of 23°C/73°F (higher temperatures reduce working time, lower temperatures extend working time).

**9.3 GENERAL PRECAUTIONS/WARNINGS ON THE USE OF THE DEVICE:**

Use is not recommended on patients who are sensitive to condensation silicones. If irritation, redness or other signs of hypersensitivity occur, stop using the product and take the steps necessary to ensure the patient's safety. In order to avoid complications, it is advisable to block excessive undercut before taking the impression. Always check that the material has set completely before removing the impression from the patient's mouth. If the impression has been taken, check its integrity and make sure there are no material residues in the patient's mouth. When performing the double step impression technique, wash the first impression and dry it thoroughly to improve adhesion between the two viscosities. As silicones leave indelible residue, avoid contact with clothing.

**9.4 PRELIMINARY OPERATIONS:**

1. Read the instructions for use.

2. Use protective gloves, face mask, protective goggles and suitable clothing.

3. Select the impression tray with retention rims.

4. Then choose the impression technique to be used.

**9.5 IMPRESSION TAKING PROCEDURE**

Two-phase technique (double step impression):

1. Prepare the high-viscosity material device to be used for the first impression. See paragraph 9.6 for details on how to prepare the device.

2. Mix high-viscosity material observing the mixing times (see paragraph 10, technical data table).

3. Place a suitable amount of the high-viscosity material on the impression tray (fig. 1A).

4. Take the first impression with the high-viscosity material tray inside the patient's mouth within the clinical working time (see paragraph 10, technical data table) (fig. 1B).

5. Remove the impression from the patient's mouth when the setting time has been reached (see paragraph 10, technical data table) (fig. 2A).

6. Wash the first impression and dry it thoroughly.

7. Prepare the low viscosity material device to be used for the second impression. See paragraph 9.6 for details on how to prepare the device.

8. Use protective gloves, face mask, protective goggles and suitable clothing.

9. Select the low-viscosity impression material, observing the mixing times (see paragraph 10, technical data table).

10. Apply a suitable amount of low-viscosity material where necessary (preparations, first impression, etc.) (fig. 3A and 4A) and place the reloaded impression tray back inside the patient's mouth to take the second impression (first working time) (see paragraph 10, technical data table) (fig. 5A).

11. Remove the impression from the patient's mouth when the setting time has been reached (see paragraph 10, technical data table) (fig. 5B).

12. Disinfect as shown in paragraphs 9.7 and 9.8. To avoid compromising the performance of the products, it is essential to respect the timing indicated in the technical data table for each phase.

**9.6 PREPARING THE DEVICES SPECIFIC WARNINGS:**

Paste Hardener

In order to reduce the risk of cross-contamination, always use new, uncontaminated gloves when handling tubes and accessories. Use with Durosil L and Durosil S. The Paste Hardener catalyst causes skin irritation. Use protective goggles, face mask, and suitable clothing. Avoid direct contact between the catalyst and skin. If irritation, redness or other signs of hypersensitivity occur, wash off with soap and water and seek immediate medical advice. In the event of accidental contact with the skin, wash thoroughly with soap and water. If irritation, redness or other signs of hypersensitivity occur, during the use of Paste Hardener, stop using the product and take the steps necessary to ensure the patient's safety. Before opening the tube of catalyst, it is recommended that you close the tube of catalyst immediately after extruding the material to avoid the nozzle getting clogged. Avoid contact with clothing because the catalyst leaves indelible stains.

**Durosil L - PASTE HARDENER:**

1. Remove any salts and lids.

2. Use the spoon provided to measure out the base. The measuring spoon should be filled flush with the surface (fig. 1).

3. Press the edge of the measuring spoon against the material (fig. 1B). Repeat the procedure for each spoonful measured (fig. 1B). Replace the cap on the tube of base.

4. Open the tube of Paste Hardener catalyst. The catalyst can be supplied separately from the bases of the materials it is combined with for mixing.

5. Spread two parallel strips of Paste Hardener catalyst of the same length as the measuring spoon, taking care to reach the inner edges of the same (equal to 4 cm) (fig. 1C). Repeat the procedure for each spoonful measured (fig. 1B).

7. Replace the cap on the tube of Paste Hardener catalyst immediately.

8. Mix the material with the fingers, by folding it back on itself repeatedly, to achieve an even colour with no streaks (fig. 1C). Observe the mixing time indicated in the technical data table.

9. Proceed in accordance with the chosen technique.

10. After use, disinfect the impression as described in the disinfection procedure (see paragraphs 9.7 and 9.8) and disinfect the spatula following the instructions provided by the manufacturer.

**Durosil S - PASTE HARDENER**

1. Remove the base cap. Squeeze a strip of base onto a mixing block or a clean surface. Close the base tube immediately after dispensing the material.

2. Open the tube of Paste Hardener catalyst. The catalyst can be supplied separately from the bases of the materials it is combined with for mixing.

3. Spread a strip of Paste Hardener catalyst the same length as the strip of base onto the mixing block or clean surface. The length of base to catalyst ratio must be 1:1 (fig. 2). Repeat the procedure for each strip of base dispensed (fig. 2B).

4. Replace the cap on the tube of Paste Hardener catalyst immediately.

5. Mix the material with the fingers, by folding it back on itself repeatedly, to obtain an even colour without any streaks (fig. 2C). Observe the mixing time indicated in the technical data table.

6. Proceed in accordance with the chosen technique.

7. After use, disinfect the impression as described in the disinfection procedure (see paragraphs 9.7 and 9.8) and disinfect the spatula following the instructions provided by the manufacturer.

**9.6 DISINFECTING AND CLEANING THE DEVICE: GENERAL PRECAUTIONS / WARNINGS**

Use clear, appropriately disinfected/sterilized accessories. The impressions must be disinfected using a disinfectant intended specifically for Condensation Silicones. Use of an unsuitable disinfectant or of the right disinfectant but for too long may compromise the impression.

**9.7 DISINFECTING AND CLEANING THE DEVICE: GENERAL STEP-BY-STEP INSTRUCTIONS**

1. Rinse the impression thoroughly to remove any residue and traces of saliva (fig. 1C).

2. Remove any excess water.

3. Disinfect the impression using a disinfectant intended specifically for Condensation Silicones (Polysiloxanes), dipping the impression in the solution or spraying it directly if using a spray disinfectant (fig. 2C). Follow the dosing instructions of the manufacturer of the disinfectant. For contact time: Quaternary ammonium salt-based disinfectants and mixtures of alcohol and surface tension reducers have been tested. Please refer to the instructions for use for each specific disinfectant you wish to use, to check compatibility and effectiveness.

**10. TECHNICAL DATA**

The times for clinical use and technical data are shown in Table 1.

	Durosil L	Durosil S
Type 0 Putty Consistency	Type 3 Light Body Consistency	
Mixing time*	30°	30°
Minimum processing time* (ISO)	1°	1°
Clinical working time* (including mixing time)	115°	130°
Time in mouth**	3'30"	3'30"
Setting time	4'45"	5'00"
Elastic recovery	98.5%	98.5%
Shore A-hardness 1 hour	70	30
Mixing ratio	18 g/0.31 g	6.5 g/0.65 g

\* The times are intended from the start of mixing to 23°C/73°F and 50% of relative humidity. Intensity: high. Temperatures and excesses of time.

\*\* The time is intended to be 35°C/95°F. Use of larger amounts of Paste Hardener will shorten this time, use of smaller amounts will extend it.

**11. STORAGE AND STABILITY**

The devices should be used within a professional dental environment (legally-certified public or private health facilities). Use the product at an ambient temperature of 23°C/73°F (higher temperatures reduce working time, lower temperatures increase working time). Do not use after the time indicated on the packaging. Store the product at a temperature between 5°C/41°F and 27°C/80°F. Do not store the product in direct sunlight. Store the Paste Hardener catalyst in a dry place at a temperature of between 5°C/41°F and 27°C/80°F. Do not store the product in the direct sunlight and protect from possible microbial contamination.

**12. IMPRESSION CASTING**

After completing disinfection, dry the impression before casting the gypsum. It is advisable to cast the impression within 72 hours after the disinfection procedure.

**13. CONSERVATION OF THE IMPRESSION** Store the impressions at ambient temperature, in a dry place away from direct sunlight.

**14. IMPORTANT REMARKS**

Information provided in any way, even during demonstrations, does not invalidate the instructions for use.

Operation are required to check that the product is suitable for the application envisaged. The manufacturer cannot be held responsible for damage, including to third parties, deriving from failure to follow instructions or from unsatisfactory use for an application. The manufacturer's liability is, in any case, limited to the value of the products supplied. Report any serious incident involving the medical device to the manufacturer and to the relevant authorities.

## DE) DUROSIL L - DUROSIL S - PASTE HARDENER®

C Silikones  
Gebrauchsanweisung1. VERWENDUNGSZWECK  
Zahnformungsmaterial aus C-Silikon

**2. PRODUCT DESCRIPTION**  
Durosil L: Hochviskose kondensationspolysiloxane (Basis) empfohlen für die Zweiphasentechnik (doppelte Abdrucknahme) in Kombination mit Durosil S.

Durosil S: Low-viscosity condensation polysiloxane (base) recommended for the two-phase technique (double step impression) in combination with Durosil L.

Paste Hardener: Catalyst for condensation polysiloxanes. To be used for the polymerisation of Durosil L and Durosil S bases only. The Paste Hardener catalyst can be supplied separately from the bases of the materials it is combined with for mixing.

## 3. VERPACKUNGSFORMEN

• 900 ml Kap. des Basisprodukts Durosil L • 140 ml tube des Basisprodukts Durosil S • 60 ml tube of Paste Hardener

## 4. ZUSAMMENSETZUNG

Durosil L: Polydimethylsiloxane (Hydroxymethyl), Siliciumdioxid-Füllstoffe, anorganische Aluminium-Füllstoffe, Kohlenwasserstoffe, Pigmente, Fruchtgeschmack.

Durosil S: Polydimethylsiloxane (Hydroxymethyl), Kieselgur-Füllstoffe, Pigmente, Aromaten (Fruchtmix).


Paste Hardener: Organogel-Katalysator für condensation polysiloxane.

## 5. GEBRAUCHSANWEISUNG

Die Vorschritte sind für den Einsatz im zahnärztlichen Bereich durch geschultes und qualifiziertes Fachpersonal zur Abrücknahme bei Zahnpatienten ab 3 Jahren vorgesehen.

Vorbereiten der Geräte: Schritt-für-Schritt-Anleitung

Durosil L & Paste Hardener:



Durosil S & Paste Hardener:

